

RUSH

RUSH

RUSH

G & L Laboratories

246 Arlington St., Quincy, MA 02170
Tel: (617) 328-3663 Fax: (617) 472-0706

Lab. ID #
113349

To: Attn: Cathleen Liberty
Franklin Health Dept.
355 E. Central St.
Franklin, MA 02038

Received Date/Time: 7/24/24, 5:55pm
Analyzed Date/Time: 7/24/24, 6:15p
Reported Date: 7/25/24
Collected by: Client
Collected Date: 7/24/24

ANALYTICAL RESULTS

Sample#	Sampling Location	Beach Location ID#	Sampling Time	E. coli (CFU/100mL)	GeoMean
1	Chilson Beach	3410	10:00AM ✓	28	
Maximum Contaminant Level				235	126
Method of Reference				EPA 1603	

Laboratory QC Results

QC	Unit	E. coli
Duplicate (D1)	100mL	<5
Duplicate (D2)	100mL	<5
Dilution Blk.(Beginning)	100mL	Negative
Dilution Blk.(Ending)	100mL	Negative
Positive Control	100mL	Positive

Notes:
T.C. = 8.1 °C

Approved by:  7/25/24
Lab. Director Date